

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017052.D
 Acq On : 30 Jun 2020 17:56
 Operator : JC/SP
 Sample : L3123-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Manual Integrations
 APPROVED

MMDadoda
 7/6/2020 4:45:56 PM

Quant Time: Jul 01 09:31:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	233869	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	397628	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	413249	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	211854	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	151472	52.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.08%	
35) Dibromofluoromethane	5.47	113	130867	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
50) Toluene-d8	8.70	98	491315	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
62) 4-Bromofluorobenzene	11.12	95	201765	55.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.98%	
Target Compounds						
4) Vinyl Chloride	1.39	62	10198	4.013	ug/l	99
12) 1,1-Dichloroethene	2.36	96	44794	19.359	ug/l	100
16) Acetone	2.42	43	7317	6.784	ug/l	93
24) 1,1-Dichloroethane	3.67	63	25457	5.653	ug/l	98
27) cis-1,2-Dichloroethene	4.57	96	10936	3.759	ug/l	87
40) Benzene	6.12	78	4581	0.415	ug/l	99
42) 1,2-Dichloroethane	6.17	62	2142m	0.537	ug/l	
49) 1,4-Dioxane	7.72	88	1693	24.826	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017052.D
Acq On : 30 Jun 2020 17:56
Operator : JC/SP
Sample : L3123-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
MW-3

Manual Integrations
APPROVED

MMDadoda
7/6/2020 4:45:56 PM

Quant Time: Jul 01 09:31:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

