

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017055.D
 Acq On : 30 Jun 2020 19:04
 Operator : JC/SP
 Sample : L3152-01 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40680

Quant Time: Jul 01 09:37:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	226208	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	389489	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	397967	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	208120	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	152564	54.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.42%	
35) Dibromofluoromethane	5.46	113	130285	52.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.76%	
50) Toluene-d8	8.70	98	472096	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
62) 4-Bromofluorobenzene	11.12	95	196909	55.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.58%	
Target Compounds						
11) Tert butyl alcohol	3.01	59	25278	47.515	ug/l	97
16) Acetone	2.42	43	1118611	1072.264	ug/l	98
20) Methylene Chloride	2.84	84	5315	1.462	ug/l #	93
40) Benzene	6.11	78	3815	0.353	ug/l	92
52) Toluene	8.76	92	50686	7.506	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017055.D
Acq On : 30 Jun 2020 19:04
Operator : JC/SP
Sample : L3152-01 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40680

Quant Time: Jul 01 09:37:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

