

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017068.D
 Acq On : 01 Jul 2020 01:07
 Operator : JC/SP
 Sample : PB129852ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129852ZHE#04

Quant Time: Jul 01 08:46:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	229936	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	399064	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	430636	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	218154	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	150532	53.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.20%	
35) Dibromofluoromethane	5.47	113	132772	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
50) Toluene-d8	8.70	98	502115	53.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.58%	
62) 4-Bromofluorobenzene	11.12	95	211282	57.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.80%	
Target Compounds						
16) Acetone	2.42	43	13731	12.949	ug/l	94
18) Methyl Acetate	2.75	43	3310	1.196	ug/l	98
20) Methylene Chloride	2.84	84	5460	1.486	ug/l #	91
43) Isopropyl Acetate	6.42	43	69357	10.907	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017068.D
Acq On : 01 Jul 2020 01:07
Operator : JC/SP
Sample : PB129852ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129852ZHE#04

Quant Time: Jul 01 08:46:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

