

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017069.D
 Acq On : 01 Jul 2020 01:30
 Operator : JC/SP
 Sample : PB129852ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129852ZHE#05

Quant Time: Jul 01 08:48:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	225948	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	392122	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	413359	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	215538	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	150341	53.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.94%	
35) Dibromofluoromethane	5.47	113	129045	52.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.04%	
50) Toluene-d8	8.70	98	493943	53.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.70%	
62) 4-Bromofluorobenzene	11.12	95	207800	57.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.92%	
Target Compounds						
16) Acetone	2.42	43	15111	14.502	ug/l	98
18) Methyl Acetate	2.76	43	3617	1.330	ug/l	96
20) Methylene Chloride	2.84	84	6289	1.878	ug/l #	91
43) Isopropyl Acetate	6.42	43	76636	12.265	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017069.D
Acq On : 01 Jul 2020 01:30
Operator : JC/SP
Sample : PB129852ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129852ZHE#05

Quant Time: Jul 01 08:48:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

