

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017071.D
 Acq On : 01 Jul 2020 02:15
 Operator : JC/SP
 Sample : PB129852ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129852ZHE#07

Quant Time: Jul 01 08:51:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	224102	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	388881	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	417942	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	212379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	147657	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
35) Dibromofluoromethane	5.46	113	129403	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
50) Toluene-d8	8.70	98	488027	53.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.30%	
62) 4-Bromofluorobenzene	11.12	95	206245	58.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.00%	
Target Compounds						
16) Acetone	2.42	43	15071	14.582	ug/l	92
18) Methyl Acetate	2.75	43	3056	1.133	ug/l #	84
20) Methylene Chloride	2.84	84	6226	1.873	ug/l	98
43) Isopropyl Acetate	6.42	43	78327	12.640	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017071.D
Acq On : 01 Jul 2020 02:15
Operator : JC/SP
Sample : PB129852ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129852ZHE#07

Quant Time: Jul 01 08:51:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

