

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017073.D
 Acq On : 01 Jul 2020 03:01
 Operator : JC/SP
 Sample : PB129852ZHE#09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129852ZHE#09

Quant Time: Jul 01 08:56:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	223551	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	388143	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	413296	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	212215	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	151492	54.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.94%	
35) Dibromofluoromethane	5.47	113	129212	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
50) Toluene-d8	8.70	98	490780	54.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.12%	
62) 4-Bromofluorobenzene	11.12	95	203522	57.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.68%	
Target Compounds						
16) Acetone	2.42	43	10625	10.306	ug/l	97
18) Methyl Acetate	2.75	43	2827	1.050	ug/l	93
20) Methylene Chloride	2.84	84	12774	4.693	ug/l	94
43) Isopropyl Acetate	6.42	43	60973	9.858	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017073.D
Acq On : 01 Jul 2020 03:01
Operator : JC/SP
Sample : PB129852ZHE#09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129852ZHE#09

Quant Time: Jul 01 08:56:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

