

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017080.D
 Acq On : 01 Jul 2020 05:40
 Operator : JC/SP
 Sample : PB129852ZHE#16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129852ZHE#16

Quant Time: Jul 01 09:06:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	213460	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	373418	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	407767	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	205072	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	144328	54.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.70%	
35) Dibromofluoromethane	5.47	113	127250	53.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.74%	
50) Toluene-d8	8.70	98	471049	53.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.86%	
62) 4-Bromofluorobenzene	11.12	95	199188	58.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.68%	
Target Compounds						
16) Acetone	2.42	43	13642	13.858	ug/l	99
18) Methyl Acetate	2.75	43	3569	1.389	ug/l	95
20) Methylene Chloride	2.84	84	5137	1.516	ug/l	85
43) Isopropyl Acetate	6.42	43	64219	10.793	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017080.D
Acq On : 01 Jul 2020 05:40
Operator : JC/SP
Sample : PB129852ZHE#16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129852ZHE#16

Quant Time: Jul 01 09:06:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

