

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017081.D
 Acq On : 01 Jul 2020 06:03
 Operator : JC/SP
 Sample : PB129852ZHE#17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129852ZHE#17

Quant Time: Jul 01 09:08:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	218457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	378844	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	404544	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	208782	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	146361	54.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.70%	
35) Dibromofluoromethane	5.47	113	127340	53.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.28%	
50) Toluene-d8	8.70	98	477948	53.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.86%	
62) 4-Bromofluorobenzene	11.12	95	195857	56.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.08%	
Target Compounds						
16) Acetone	2.42	43	13957	13.853	ug/l	98
18) Methyl Acetate	2.76	43	3632	1.381	ug/l	99
20) Methylene Chloride	2.84	84	5988	1.838	ug/l	97
43) Isopropyl Acetate	6.42	43	68739	11.387	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017081.D
Acq On : 01 Jul 2020 06:03
Operator : JC/SP
Sample : PB129852ZHE#17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129852ZHE#17

Quant Time: Jul 01 09:08:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

