

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX063020\
 Data File : VX017083.D
 Acq On : 01 Jul 2020 06:48
 Operator : JC/SP
 Sample : PB129852ZHE#19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129852ZHE#19

Quant Time: Jul 01 09:11:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	232373	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	401166	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	437729	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	222511	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	156808	54.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.48%	
35) Dibromofluoromethane	5.46	113	136344	53.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.46%	
50) Toluene-d8	8.70	98	514434	54.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.64%	
62) 4-Bromofluorobenzene	11.12	95	212762	58.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.00%	
Target Compounds						
16) Acetone	2.42	43	15374	14.346	ug/l	99
18) Methyl Acetate	2.75	43	3807	1.361	ug/l #	86
20) Methylene Chloride	2.84	84	6183	1.761	ug/l	95
43) Isopropyl Acetate	6.42	43	73597	11.513	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX063020\
Data File : VX017083.D
Acq On : 01 Jul 2020 06:48
Operator : JC/SP
Sample : PB129852ZHE#19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129852ZHE#19

Quant Time: Jul 01 09:11:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

