

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX063020\
 Data File : VX017032.D
 Acq On : 30 Jun 2020 09:49
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 01 08:07:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.484	0.503	-3.9	107	0.00
3 P	Chloromethane	0.586	0.550	6.1	100	0.00
4 C	Vinyl Chloride	0.543	0.524	3.5#	102	0.00
5 T	Bromomethane	0.287	0.299	-4.2	108	0.00
6 T	Chloroethane	0.321	0.316	1.6	105	0.00
7 T	Trichlorofluoromethane	0.803	0.832	-3.6	110	0.00
8 T	Diethyl Ether	0.293	0.285	2.7	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.482	0.461	4.4	98	0.00
10 T	Methyl Iodide	0.621	0.629	-1.3	98	0.00
11 T	Tert butyl alcohol	0.118	0.105	11.0	91	0.00
12 CM	1,1-Dichloroethene	0.495	0.442	10.7#	93	0.00
13 T	Acrolein	0.070	0.063	10.0	93	0.00
14 T	Allyl chloride	0.914	0.860	5.9	99	0.00
15 T	Acrylonitrile	0.283	0.265	6.4	96	0.00
16 T	Acetone	0.231	0.190	17.7	86	0.00
17 T	Carbon Disulfide	1.515	1.426	5.9	102	0.00
18 T	Methyl Acetate	0.602	0.559	7.1	96	0.00
19 T	Methyl tert-butyl Ether	1.660	1.643	1.0	101	0.00
20 T	Methylene Chloride	0.589	0.547	7.1	100	0.00
21 T	trans-1,2-Dichloroethene	0.549	0.544	0.9	103	0.00
22 T	Diisopropyl ether	1.692	1.709	-1.0	102	0.00
23 T	Vinyl Acetate	1.454	1.489	-2.4	100	0.00
24 P	1,1-Dichloroethane	0.963	0.940	2.4	101	0.00
25 T	2-Butanone	0.384	0.350	8.9	92	0.00
26 T	2,2-Dichloropropane	0.855	0.857	-0.2	105	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.590	5.1	99	0.00
28 T	Bromochloromethane	0.468	0.451	3.6	101	0.00
29 T	Tetrahydrofuran	0.254	0.238	6.3	93	0.00
30 C	Chloroform	0.987	0.974	1.3#	102	0.00
31 T	Cyclohexane	0.832	0.855	-2.8	103	0.00
32 T	1,1,1-Trichloroethane	0.909	0.883	2.9	101	0.00
33 S	1,2-Dichloroethane-d4	0.616	0.569	7.6	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.316	0.318	-0.6	102	0.00
36 T	1,1-Dichloropropene	0.473	0.495	-4.7	103	0.00
37 T	Ethyl Acetate	0.479	0.479	0.0	93	0.00
38 T	Carbon Tetrachloride	0.520	0.562	-8.1	105	0.00
39 T	Methylcyclohexane	0.526	0.589	-12.0	107	0.00
40 TM	Benzene	1.388	1.439	-3.7	101	0.00
41 T	Methacrylonitrile	0.270	0.275	-1.9	99	0.00
42 TM	1,2-Dichloroethane	0.502	0.515	-2.6	101	0.00
43 T	Isopropyl Acetate	0.797	0.791	0.8	96	0.00
44 TM	Trichloroethene	0.418	0.419	-0.2	102	0.00
45 C	1,2-Dichloropropane	0.357	0.375	-5.0#	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX063020\
 Data File : VX017032.D
 Acq On : 30 Jun 2020 09:49
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 01 08:07:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.243	0.252	-3.7	101	0.00
47 T	Bromodichloromethane	0.502	0.529	-5.4	102	0.00
48 T	Methyl methacrylate	0.379	0.395	-4.2	96	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	94	0.00
50 S	Toluene-d8	1.170	1.171	-0.1	103	0.00
51 T	4-Methyl-2-Pentanone	0.475	0.487	-2.5	96	0.00
52 CM	Toluene	0.867	0.919	-6.0#	102	0.00
53 T	t-1,3-Dichloropropene	0.554	0.579	-4.5	100	0.00
54 T	cis-1,3-Dichloropropene	0.593	0.628	-5.9	101	0.00
55 T	1,1,2-Trichloroethane	0.354	0.375	-5.9	104	0.00
56 T	Ethyl methacrylate	0.504	0.544	-7.9	99	0.00
57 T	1,3-Dichloropropane	0.599	0.618	-3.2	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.251	0.297	-18.3	105	0.00
59 T	2-Hexanone	0.347	0.370	-6.6	98	0.00
60 T	Dibromochloromethane	0.411	0.439	-6.8	103	0.00
61 T	1,2-Dibromoethane	0.381	0.398	-4.5	101	0.00
62 S	4-Bromofluorobenzene	0.457	0.457	0.0	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.404	0.430	-6.4	104	0.00
65 PM	Chlorobenzene	1.032	1.085	-5.1	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.398	0.419	-5.3	102	0.00
67 C	Ethyl Benzene	1.723	1.888	-9.6#	102	0.00
68 T	m/p-Xylenes	0.662	0.741	-11.9	103	0.00
69 T	o-Xylene	0.630	0.699	-11.0	103	0.00
70 T	Styrene	1.072	1.227	-14.5	103	0.00
71 P	Bromoform	0.315	0.348	-10.5	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.275	3.542	-8.2	104	0.00
74 T	N-amyl acetate	1.397	1.456	-4.2	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.030	1.032	-0.2	101	0.00
76 T	1,2,3-Trichloropropane	0.983	1.008	-2.5	99	0.00
77 T	Bromobenzene	0.900	0.916	-1.8	102	0.00
78 T	n-propylbenzene	3.704	4.091	-10.4	106	0.00
79 T	2-Chlorotoluene	2.241	2.408	-7.5	104	0.00
80 T	1,3,5-Trimethylbenzene	2.717	3.025	-11.3	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.380	0.395	-3.9	99	0.00
82 T	4-Chlorotoluene	2.679	2.871	-7.2	105	0.00
83 T	tert-Butylbenzene	2.667	2.902	-8.8	103	0.00
84 T	1,2,4-Trimethylbenzene	2.756	3.056	-10.9	104	0.00
85 T	sec-Butylbenzene	3.135	3.480	-11.0	105	0.00
86 T	p-Isopropyltoluene	2.918	3.239	-11.0	103	0.00
87 T	1,3-Dichlorobenzene	1.635	1.662	-1.7	106	0.00
88 T	1,4-Dichlorobenzene	1.675	1.683	-0.5	105	0.00
89 T	n-Butylbenzene	2.606	2.825	-8.4	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX063020\
Data File : VX017032.D
Acq On : 30 Jun 2020 09:49
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 01 08:07:44 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.575	0.613	-6.6	104	0.00
91 T	1,2-Dichlorobenzene	1.528	1.546	-1.2	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.246	0.231	6.1	97	0.00
93 T	1,2,4-Trichlorobenzene	1.043	1.068	-2.4	103	0.00
94 T	Hexachlorobutadiene	0.422	0.438	-3.8	110	0.00
95 T	Naphthalene	3.171	3.364	-6.1	99	0.00
96 T	1,2,3-Trichlorobenzene	1.009	1.050	-4.1	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6