


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX063020\
 Data File : VX017059.D
 Acq On : 30 Jun 2020 20:57
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Title : SW846 8260
 Last Update : Mon Jun 29 17:10:38 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	12675	PASS
75	95	30	60	49.6	36413	PASS
95	95	100	100	100.0	73429	PASS
96	95	5	9	6.9	5075	PASS
173	174	0.00	2	1.3	741	PASS
174	95	50	100	79.8	58608	PASS
175	174	5	9	7.6	4453	PASS
176	174	95	101	99.8	58520	PASS
177	176	5	9	6.7	3928	PASS