

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023023.D
 Acq On : 30 Jun 2021 10:55
 Operator : JC/MD
 Sample : VX0630MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0630MBL01

Quant Time: Jul 01 03:24:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	59476	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	121399	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105516	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	53653	49.708	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.420%
35) Dibromofluoromethane	5.391	113	38060	49.370	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.740%
50) Toluene-d8	8.653	98	148159	49.037	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.080%
62) 4-Bromofluorobenzene	11.085	95	49818	44.181	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023023.D
Acq On : 30 Jun 2021 10:55
Operator : JC/MD
Sample : VX0630MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0630MBL01

Quant Time: Jul 01 03:24:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

