

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023024.D
 Acq On : 30 Jun 2021 11:18
 Operator : JC/MD
 Sample : VX0630WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0630WBL01

Quant Time: Jul 01 03:24:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	57085	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.769	114	115638	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	103109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37617	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	59143	57.089	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.180%
35) Dibromofluoromethane	5.397	113	40462	55.101	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.200%
50) Toluene-d8	8.653	98	157775	54.822	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.640%
62) 4-Bromofluorobenzene	11.085	95	54155	50.420	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023024.D
Acq On : 30 Jun 2021 11:18
Operator : JC/MD
Sample : VX0630WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0630WBL01

Quant Time: Jul 01 03:24:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

