

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023027.D
 Acq On : 30 Jun 2021 12:35
 Operator : JC/MD
 Sample : M2875-12DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6DL

Quant Time: Jul 01 03:25:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound			R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards								
1) Pentafluorobenzene			5.562	168	62176	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene			6.769	114	123681	50.000	ug/l	0.00
63) Chlorobenzene-d5			10.061	117	107663	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4			12.024	152	40888	50.000	ug/l	0.00
System Monitoring Compounds								
33) 1,2-Dichloroethane-d4			5.964	65	56517	50.088	ug/l	0.00
Spiked Amount 50.000			Range 78 - 117		Recovery	= 100.180%		
35) Dibromofluoromethane			5.391	113	38216	48.658	ug/l	0.00
Spiked Amount 50.000			Range 75 - 124		Recovery	= 97.320%		
50) Toluene-d8			8.653	98	153130	49.747	ug/l	0.00
Spiked Amount 50.000			Range 92 - 112		Recovery	= 99.500%		
62) 4-Bromofluorobenzene			11.085	95	52622	45.806	ug/l	0.00
Spiked Amount 50.000			Range 83 - 123		Recovery	= 91.620%		
Target Compounds								
								Qvalue
19) Methyl tert-butyl Ether			3.123	73	2326	0.804	ug/l	# 1
39) Methylcyclohexane			7.379	83	2126	1.387	ug/l	# 81
40) Benzene			6.044	78	55300	14.598	ug/l	100
52) Toluene			8.720	92	16235	7.090	ug/l	99
67) Ethyl Benzene			10.195	91	55289	13.243	ug/l	99
68) m/p-Xylenes			10.305	106	77580	50.036	ug/l	94
69) o-Xylene			10.646	106	23727	15.723	ug/l	93
73) Isopropylbenzene			10.963	105	3214	0.929	ug/l	89
78) n-propylbenzene			11.305	91	5540	1.323	ug/l	97
80) 1,3,5-Trimethylbenzene			11.457	105	11076	3.899	ug/l	94
84) 1,2,4-Trimethylbenzene			11.756	105	46265	16.181	ug/l	94
95) Naphthalene			13.780	128	9291	4.603	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023027.D
Acq On : 30 Jun 2021 12:35
Operator : JC/MD
Sample : M2875-12DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6DL

Quant Time: Jul 01 03:25:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

