

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023029.D
 Acq On : 30 Jun 2021 13:22
 Operator : JC/MD
 Sample : M2872-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 01 03:25:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	58501	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	118177	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38889	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53801	50.676	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.360%	
35) Dibromofluoromethane	5.397	113	36954	49.242	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.480%	
50) Toluene-d8	8.653	98	148156	50.373	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.740%	
62) 4-Bromofluorobenzene	11.085	95	51389	46.816	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.640%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023029.D
Acq On : 30 Jun 2021 13:22
Operator : JC/MD
Sample : M2872-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 01 03:25:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

