

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX063021\  
 Data File : VX023031.D  
 Acq On : 30 Jun 2021 14:09  
 Operator : JC/MD  
 Sample : M2872-04  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Jul 01 03:26:08 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X062321W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 23 19:12:26 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.562  | 168   | 58836    | 50.000   | ug/l       | # 0.00   |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 122077   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 106407   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 38219    | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 61015    | 57.144   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 114.280% |          |
| 35) Dibromofluoromethane    | 5.391  | 113   | 37617    | 48.524   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 97.040%  |          |
| 50) Toluene-d8              | 8.653  | 98    | 150266   | 49.458   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 98.920%  |          |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 50090    | 44.175   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 88.360%  |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX063021\  
Data File : VX023031.D  
Acq On : 30 Jun 2021 14:09  
Operator : JC/MD  
Sample : M2872-04  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Jul 01 03:26:08 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X062321W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jun 23 19:12:26 2021  
Response via : Initial Calibration

