

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023033.D
 Acq On : 30 Jun 2021 14:56
 Operator : JC/MD
 Sample : M2897-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jul 01 03:26:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	51654	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	108110	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35221	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56203	59.956	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	119.920%#
35) Dibromofluoromethane	5.391	113	34593	50.389	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.780%
50) Toluene-d8	8.653	98	134179	49.869	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.740%
62) 4-Bromofluorobenzene	11.085	95	45310	45.122	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.240%
Target Compounds						
20) Methylene Chloride	2.794	84	402	0.856	ug/l	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023033.D
Acq On : 30 Jun 2021 14:56
Operator : JC/MD
Sample : M2897-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jul 01 03:26:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

