

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023034.D
 Acq On : 30 Jun 2021 15:20
 Operator : JC/MD
 Sample : M2892-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Jul 01 03:26:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	50928	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	109092	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	95210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	33805	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	50581	54.727	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.460%
35) Dibromofluoromethane	5.397	113	34215	49.389	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.780%
50) Toluene-d8	8.653	98	132114	48.660	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.320%
62) 4-Bromofluorobenzene	11.085	95	44948	44.359	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.720%
Target Compounds						
16) Acetone	2.386	43	3415	8.335	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023034.D
Acq On : 30 Jun 2021 15:20
Operator : JC/MD
Sample : M2892-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Jul 01 03:26:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

