

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023037.D
 Acq On : 30 Jun 2021 16:30
 Operator : JC/MD
 Sample : M2892-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6

Quant Time: Jul 01 03:26:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53442	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	110197	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	95534	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	33229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51466	53.065	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.140%
35) Dibromofluoromethane	5.397	113	34032	48.633	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.260%
50) Toluene-d8	8.653	98	135351	49.352	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.700%
62) 4-Bromofluorobenzene	11.085	95	43725	42.719	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.440%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1064	2.475	ug/l	90
19) Methyl tert-butyl Ether	3.123	73	23030	9.263	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023037.D
Acq On : 30 Jun 2021 16:30
Operator : JC/MD
Sample : M2892-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Jul 01 03:26:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

