

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023040.D
 Acq On : 30 Jun 2021 17:41
 Operator : JC/MD
 Sample : M2892-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Jul 01 03:27:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	54200	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	113002	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35472	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51659	52.519	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	105.040%	
35) Dibromofluoromethane	5.397	113	35219	49.080	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.160%	
50) Toluene-d8	8.653	98	140423	49.931	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	99.860%	
62) 4-Bromofluorobenzene	11.085	95	46778	44.567	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	89.140%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	260521	1270.839	ug/l	98
16) Acetone	2.380	43	1568	3.596	ug/l #	83
19) Methyl tert-butyl Ether	3.123	73	1186565	470.570	ug/l	97
22) Diisopropyl ether	3.770	45	20797	7.484	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023040.D
Acq On : 30 Jun 2021 17:41
Operator : JC/MD
Sample : M2892-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Jul 01 03:27:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

