

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023043.D
 Acq On : 30 Jun 2021 18:52
 Operator : JC/MD
 Sample : M2898-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-062921

Quant Time: Jul 01 03:27:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	58933	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	117995	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107544	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46932	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53375	49.906	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.820%
35) Dibromofluoromethane	5.397	113	37052	49.449	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.900%
50) Toluene-d8	8.653	98	149179	50.799	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.600%
62) 4-Bromofluorobenzene	11.085	95	56142	51.225	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023043.D
Acq On : 30 Jun 2021 18:52
Operator : JC/MD
Sample : M2898-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-062921

Quant Time: Jul 01 03:27:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

