

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023055.D
 Acq On : 01 Jul 2021 00:44
 Operator : JC/MD
 Sample : PB137406TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137406TB

Quant Time: Jul 01 03:30:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56326	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	115318	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105678	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38995	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53381	52.222	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.440%
35) Dibromofluoromethane	5.391	113	36106	49.305	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.620%
50) Toluene-d8	8.653	98	146022	50.879	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.760%
62) 4-Bromofluorobenzene	11.085	95	51273	47.869	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.740%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	5850	12.909	ug/l	98
20) Methylene Chloride	2.794	84	6671	7.739	ug/l #	77
43) Isopropyl Acetate	6.348	43	11623	5.095	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023055.D
Acq On : 01 Jul 2021 00:44
Operator : JC/MD
Sample : PB137406TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137406TB

Quant Time: Jul 01 03:30:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

