

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023056.D
 Acq On : 01 Jul 2021 01:08
 Operator : JC/MD
 Sample : PB137429TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137429TB

Quant Time: Jul 01 03:30:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	52300	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	6.769	114	112045	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	100587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37970	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50273	52.967	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.940%
35) Dibromofluoromethane	5.403	113	33696	47.358	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.720%
50) Toluene-d8	8.653	98	137232	49.213	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.420%
62) 4-Bromofluorobenzene	11.085	95	48296	46.407	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.820%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6672	15.856	ug/l	100
20) Methylene Chloride	2.794	84	8114	10.023	ug/l #	81
25) 2-Butanone	4.574	43	3699	5.653	ug/l	92
43) Isopropyl Acetate	6.348	43	11860	5.351	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023056.D
Acq On : 01 Jul 2021 01:08
Operator : JC/MD
Sample : PB137429TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137429TB

Quant Time: Jul 01 03:30:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

