

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023057.D
 Acq On : 01 Jul 2021 01:31
 Operator : JC/MD
 Sample : PB137429ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137429ZHE#01

Quant Time: Jul 01 03:30:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	54892	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	114388	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39278	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51530	51.728	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.460%		
35) Dibromofluoromethane	5.403	113	36499	50.247	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.500%		
50) Toluene-d8	8.653	98	145348	51.055	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.120%		
62) 4-Bromofluorobenzene	11.085	95	51243	48.230	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.460%		
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	6403	14.499	ug/l	99
20) Methylene Chloride	2.794	84	7748	9.152	ug/l #	78
25) 2-Butanone	4.574	43	3656	5.324	ug/l #	88
43) Isopropyl Acetate	6.354	43	11475	5.071	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023057.D
Acq On : 01 Jul 2021 01:31
Operator : JC/MD
Sample : PB137429ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137429ZHE#01

Quant Time: Jul 01 03:30:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

