

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023058.D
 Acq On : 01 Jul 2021 01:55
 Operator : JC/MD
 Sample : PB137429ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137429ZHE#02

Quant Time: Jul 01 03:30:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53436	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	109493	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	101884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36383	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51251	52.850	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.700%
35) Dibromofluoromethane	5.397	113	35341	50.828	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.660%
50) Toluene-d8	8.653	98	141940	52.087	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.180%
62) 4-Bromofluorobenzene	11.085	95	49095	48.274	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.540%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6929	16.117	ug/l	90
20) Methylene Chloride	2.800	84	8790	10.604	ug/l	92
25) 2-Butanone	4.574	43	3939	5.892	ug/l #	86
43) Isopropyl Acetate	6.348	43	11981	5.532	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023058.D
Acq On : 01 Jul 2021 01:55
Operator : JC/MD
Sample : PB137429ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137429ZHE#02

Quant Time: Jul 01 03:30:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

