

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023059.D
 Acq On : 01 Jul 2021 02:19
 Operator : JC/MD
 Sample : PB137429ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137429ZHE#03

Quant Time: Jul 01 03:30:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	52759	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	6.769	114	111022	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	102687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35746	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51030	53.297	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.600%
35) Dibromofluoromethane	5.391	113	34265	48.602	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.200%
50) Toluene-d8	8.653	98	140394	50.810	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.620%
62) 4-Bromofluorobenzene	11.085	95	48471	47.004	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.000%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6652	15.671	ug/l	97
20) Methylene Chloride	2.800	84	7579	9.308	ug/l	92
25) 2-Butanone	4.587	43	3532	5.351	ug/l #	85
43) Isopropyl Acetate	6.354	43	11648	5.304	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023059.D
Acq On : 01 Jul 2021 02:19
Operator : JC/MD
Sample : PB137429ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137429ZHE#03

Quant Time: Jul 01 03:30:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

