

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023061.D
 Acq On : 01 Jul 2021 03:06
 Operator : JC/MD
 Sample : M2883-01 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GLYCOL

Quant Time: Jul 01 07:04:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53919	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	112659	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96018	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	32314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	52532	53.685	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.380%
35) Dibromofluoromethane	5.397	113	33396	46.681	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.360%
50) Toluene-d8	8.653	98	137892	49.180	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.360%
62) 4-Bromofluorobenzene	11.085	95	44415	42.445	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.880%
Target Compounds						Qvalue
11) Tert butyl alcohol	2.971	59	27370	134.209	ug/l	96
16) Acetone	2.386	43	8428	19.428	ug/l	# 88
20) Methylene Chloride	2.800	84	637	1.107	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023061.D
Acq On : 01 Jul 2021 03:06
Operator : JC/MD
Sample : M2883-01 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GLYCOL

Quant Time: Jul 01 07:04:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

