

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023062.D
 Acq On : 01 Jul 2021 03:29
 Operator : JC/MD
 Sample : M2845-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 41234

Quant Time: Jul 01 07:04:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	48793	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	102729	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	92156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	32706	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47767	53.944	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.880%
35) Dibromofluoromethane	5.397	113	32018	49.081	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.160%
50) Toluene-d8	8.653	98	124377	48.648	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.300%
62) 4-Bromofluorobenzene	11.085	95	43447	45.533	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.060%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	3374	8.595	ug/l	94
19) Methyl tert-butyl Ether	3.123	73	1190	0.524	ug/l	# 82
20) Methylene Chloride	2.800	84	792	1.382	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023062.D
Acq On : 01 Jul 2021 03:29
Operator : JC/MD
Sample : M2845-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
41234

Quant Time: Jul 01 07:04:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

