

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023064.D
 Acq On : 01 Jul 2021 04:16
 Operator : JC/MD
 Sample : M2829-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 15B-STOCKPILE-1

Quant Time: Jul 01 07:04:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53551	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	108580	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99876	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36093	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50762	52.233	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.460%	
35) Dibromofluoromethane	5.398	113	34450	49.963	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.920%	
50) Toluene-d8	8.653	98	137898	51.030	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.060%	
62) 4-Bromofluorobenzene	11.086	95	47218	46.819	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.640%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6112	14.186	ug/l	# 88
20) Methylene Chloride	2.794	84	18465	21.820	ug/l	# 85
43) Isopropyl Acetate	6.355	43	6268	2.918	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023064.D
Acq On : 01 Jul 2021 04:16
Operator : JC/MD
Sample : M2829-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
15B-STOCKPILE-1

Quant Time: Jul 01 07:04:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

