

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023065.D
 Acq On : 01 Jul 2021 04:40
 Operator : JC/MD
 Sample : M2829-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 15B-STOCKPILE-2A

Quant Time: Jul 01 07:05:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	55364	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	118046	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	103419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38425	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	54894	54.635	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.260%
35) Dibromofluoromethane	5.397	113	37281	49.733	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.460%
50) Toluene-d8	8.653	98	149388	50.849	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.700%
62) 4-Bromofluorobenzene	11.085	95	50074	45.669	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.340%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	7693	17.271	ug/l	# 87
20) Methylene Chloride	2.794	84	23980	27.314	ug/l	# 82
43) Isopropyl Acetate	6.348	43	8488	3.635	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023065.D
Acq On : 01 Jul 2021 04:40
Operator : JC/MD
Sample : M2829-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
15B-STOCKPILE-2A

Quant Time: Jul 01 07:05:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

