

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023067.D
 Acq On : 01 Jul 2021 05:27
 Operator : JC/MD
 Sample : M2829-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 15B-STOCKPILE-3A

Quant Time: Jul 01 07:05:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	51323	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.775	114	102859	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	95800	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	32492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54012	57.990	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 115.980%	
35) Dibromofluoromethane	5.403	113	34018	52.081	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.160%	
50) Toluene-d8	8.653	98	135105	52.777	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.560%	
62) 4-Bromofluorobenzene	11.085	95	44474	46.551	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.100%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	5822	14.100	ug/l	99
20) Methylene Chloride	2.794	84	24589	30.174	ug/l #	84
43) Isopropyl Acetate	6.354	43	8987	4.417	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023067.D
Acq On : 01 Jul 2021 05:27
Operator : JC/MD
Sample : M2829-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
15B-STOCKPILE-3A

Quant Time: Jul 01 07:05:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

