

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023068.D
 Acq On : 01 Jul 2021 05:50
 Operator : JC/MD
 Sample : M2829-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 15B-STOCKPILE-3B

Quant Time: Jul 01 07:05:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	52608	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	6.769	114	107625	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49438	51.782	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.560%
35) Dibromofluoromethane	5.391	113	34106	49.903	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.800%
50) Toluene-d8	8.653	98	136829	51.083	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.160%
62) 4-Bromofluorobenzene	11.085	95	46938	46.954	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.900%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	5101	12.052	ug/l	# 88
20) Methylene Chloride	2.794	84	19332	23.230	ug/l	# 84
43) Isopropyl Acetate	6.348	43	8216	3.859	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023068.D
Acq On : 01 Jul 2021 05:50
Operator : JC/MD
Sample : M2829-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
15B-STOCKPILE-3B

Quant Time: Jul 01 07:05:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

