

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023069.D
 Acq On : 01 Jul 2021 06:14
 Operator : JC/MD
 Sample : M2829-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 15B-DUP-01

Quant Time: Jul 01 07:05:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	52569	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	111547	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98846	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36307	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	50221	52.642	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.280%
35) Dibromofluoromethane	5.403	113	33543	47.354	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.700%
50) Toluene-d8	8.653	98	138857	50.018	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.040%
62) 4-Bromofluorobenzene	11.085	95	47725	46.063	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.120%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	5283	12.491	ug/l	# 81
20) Methylene Chloride	2.794	84	24017	28.791	ug/l	# 84
43) Isopropyl Acetate	6.348	43	8597	3.896	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023069.D
Acq On : 01 Jul 2021 06:14
Operator : JC/MD
Sample : M2829-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
15B-DUP-01

Quant Time: Jul 01 07:05:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

