

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023071.D
 Acq On : 01 Jul 2021 07:01
 Operator : JC/MD
 Sample : M2882-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DRIP-COMP-AUD-1121-1122

Quant Time: Jul 01 07:37:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	63112	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	125481	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	65347	57.054	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.100%
35) Dibromofluoromethane	5.397	113	35811	44.942	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.880%
50) Toluene-d8	8.653	98	152573	48.856	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.720%
62) 4-Bromofluorobenzene	11.085	95	52831	45.329	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.660%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.983	59	38142	159.786	ug/l	96
16) Acetone	2.404	43	203795	401.357	ug/l	# 54
17) Carbon Disulfide	2.502	76	2355	4.497	ug/l	# 89
25) 2-Butanone	4.587	43	2438095	3087.738	ug/l	92
40) Benzene	6.050	78	22837	5.942	ug/l	96
52) Toluene	8.726	92	51412	22.129	ug/l	98
68) m/p-Xylenes	10.305	106	6510	4.217	ug/l	93
69) o-Xylene	10.646	106	3367	2.241	ug/l	97
70) Styrene	10.659	104	4795	1.923	ug/l	94
84) 1,2,4-Trimethylbenzene	11.756	105	3487	1.289	ug/l	88
95) Naphthalene	13.780	128	16322	6.899	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023071.D
Acq On : 01 Jul 2021 07:01
Operator : JC/MD
Sample : M2882-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRIP-COMP-AUD-1121-1122

Quant Time: Jul 01 07:37:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

