

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023022.D
 Acq On : 30 Jun 2021 10:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 01 03:24:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.705	0.709	-0.6	97	0.00
3 P	Chloromethane	0.775	0.746	3.7	101	0.00
4 C	Vinyl Chloride	0.715	0.695	2.8#	97	0.00
5 T	Bromomethane	0.394	0.323	18.0	89	0.00
6 T	Chloroethane	0.442	0.426	3.6	96	0.00
7 T	Trichlorofluoromethane	1.038	1.062	-2.3	103	0.00
8 T	Diethyl Ether	0.401	0.394	1.7	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.650	0.686	-5.5	106	0.00
10 T	Methyl Iodide	0.722	0.674	6.6	93	0.00
11 T	Tert butyl alcohol	0.189	0.187	1.1	104	0.00
12 CM	1,1-Dichloroethene	0.649	0.642	1.1#	101	0.00
13 T	Acrolein	0.118	0.112	5.1	101	0.00
14 T	Allyl chloride	1.232	1.289	-4.6	106	0.00
15 T	Acrylonitrile	0.432	0.458	-6.0	104	0.00
16 T	Acetone	0.402	0.461	-14.7	123	0.00
17 T	Carbon Disulfide	1.635	1.513	7.5	97	0.00
18 T	Methyl Acetate	0.990	1.034	-4.4	109	0.00
19 T	Methyl tert-butyl Ether	2.326	2.454	-5.5	105	0.00
20 T	Methylene Chloride	0.848	0.785	7.4	103	0.00
21 T	trans-1,2-Dichloroethene	0.720	0.696	3.3	100	0.00
22 T	Diisopropyl ether	2.563	2.711	-5.8	104	0.00
23 T	Vinyl Acetate	2.150	2.372	-10.3	105	0.00
24 P	1,1-Dichloroethane	1.369	1.411	-3.1	103	0.00
25 T	2-Butanone	0.626	0.709	-13.3	113	0.00
26 T	2,2-Dichloropropane	1.039	1.138	-9.5	107	0.00
27 T	cis-1,2-Dichloroethene	0.799	0.825	-3.3	103	0.00
28 T	Bromochloromethane	0.645	0.666	-3.3	106	0.00
29 T	Tetrahydrofuran	0.402	0.426	-6.0	103	0.00
30 C	Chloroform	1.378	1.399	-1.5#	103	0.00
31 T	Cyclohexane	1.281	1.314	-2.6	104	0.00
32 T	1,1,1-Trichloroethane	1.107	1.150	-3.9	102	0.00
33 S	1,2-Dichloroethane-d4	0.907	0.855	5.7	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.318	0.303	4.7	98	0.00
36 T	1,1-Dichloropropene	0.531	0.550	-3.6	106	0.00
37 T	Ethyl Acetate	0.615	0.638	-3.7	104	0.00
38 T	Carbon Tetrachloride	0.445	0.479	-7.6	102	0.00
39 T	Methylcyclohexane	0.620	0.651	-5.0	103	0.00
40 TM	Benzene	1.531	1.583	-3.4	103	0.00
41 T	Methacrylonitrile	0.339	0.362	-6.8	109	0.00
42 TM	1,2-Dichloroethane	0.563	0.607	-7.8	104	0.00
43 T	Isopropyl Acetate	0.989	1.065	-7.7	106	0.00
44 TM	Trichloroethene	0.363	0.362	0.3	100	0.00
45 C	1,2-Dichloropropane	0.395	0.416	-5.3#	103	0.00
46 T	Dibromomethane	0.262	0.273	-4.2	105	0.00
47 T	Bromodichloromethane	0.466	0.546	-17.2	111	0.00
48 T	Methyl methacrylate	0.458	0.505	-10.3	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023022.D
 Acq On : 30 Jun 2021 10:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 01 03:24:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	95	0.00
50 S	Toluene-d8	1.244	1.161	6.7	96	0.00
51 T	4-Methyl-2-Pentanone	0.616	0.662	-7.5	101	0.00
52 CM	Toluene	0.926	0.942	-1.7#	99	0.00
53 T	t-1,3-Dichloropropene	0.528	0.588	-11.4	102	0.00
54 T	cis-1,3-Dichloropropene	0.594	0.662	-11.4	104	0.00
55 T	1,1,2-Trichloroethane	0.371	0.380	-2.4	99	0.00
56 T	Ethyl methacrylate	0.607	0.663	-9.2	101	0.00
57 T	1,3-Dichloropropane	0.675	0.684	-1.3	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.309	0.356	-15.2	106	0.00
59 T	2-Hexanone	0.467	0.530	-13.5	107	0.00
60 T	Dibromochloromethane	0.311	0.343	-10.3	102	0.00
61 T	1,2-Dibromoethane	0.377	0.394	-4.5	102	0.00
62 S	4-Bromofluorobenzene	0.464	0.437	5.8	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.351	0.359	-2.3	100	0.00
65 PM	Chlorobenzene	1.053	1.073	-1.9	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.340	0.373	-9.7	101	0.00
67 C	Ethyl Benzene	1.939	2.049	-5.7#	102	0.00
68 T	m/p-Xylenes	0.720	0.737	-2.4	100	0.00
69 T	o-Xylene	0.701	0.722	-3.0	100	0.00
70 T	Styrene	1.163	1.252	-7.7	103	0.00
71 P	Bromoform	0.194	0.218	-12.4	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	4.231	4.571	-8.0	102	0.00
74 T	N-amyl acetate	2.118	2.410	-13.8	103	0.00
75 P	1,1,2,2-Tetrachloroethane	1.453	1.521	-4.7	99	0.00
76 T	1,2,3-Trichloropropane	1.321	1.461	-10.6	105	0.00
77 T	Bromobenzene	0.880	0.923	-4.9	102	0.00
78 T	n-propylbenzene	5.119	5.714	-11.6	104	0.00
79 T	2-Chlorotoluene	3.091	3.325	-7.6	102	0.00
80 T	1,3,5-Trimethylbenzene	3.474	3.860	-11.1	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.427	0.459	-7.5	101	0.00
82 T	4-Chlorotoluene	3.572	3.898	-9.1	106	0.00
83 T	tert-Butylbenzene	3.281	3.598	-9.7	104	0.00
84 T	1,2,4-Trimethylbenzene	3.496	3.929	-12.4	103	0.00
85 T	sec-Butylbenzene	4.424	4.992	-12.8	105	0.00
86 T	p-Isopropyltoluene	3.516	3.973	-13.0	105	0.00
87 T	1,3-Dichlorobenzene	1.697	1.820	-7.2	103	0.00
88 T	1,4-Dichlorobenzene	1.775	1.817	-2.4	102	0.00
89 T	n-Butylbenzene	3.343	3.882	-16.1	106	0.00
90 T	Hexachloroethane	0.549	0.619	-12.8	103	0.00
91 T	1,2-Dichlorobenzene	1.657	1.746	-5.4	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.285	0.330	-15.8	105	0.00
93 T	1,2,4-Trichlorobenzene	0.919	1.009	-9.8	104	0.00
94 T	Hexachlorobutadiene	0.360	0.400	-11.1	114	0.00
95 T	Naphthalene	3.590	4.125	-14.9	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
Data File : VX023022.D
Acq On : 30 Jun 2021 10:23
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 01 03:24:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.925	1.047	-13.2	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6