


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063021\
 Data File : VX023050.D
 Acq On : 30 Jun 2021 21:36
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Title : SW846 8260
 Last Update : Wed Jun 23 19:12:26 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.7	5159	PASS
75	95	30	60	57.1	13598	PASS
95	95	100	100	100.0	23824	PASS
96	95	5	9	6.8	1609	PASS
173	174	0.00	2	1.4	213	PASS
174	95	50	100	62.4	14877	PASS
175	174	5	9	7.5	1121	PASS
176	174	95	101	95.5	14213	PASS
177	176	5	9	6.1	866	PASS