

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063023\
 Data File : VX036387.D
 Acq On : 30 Jun 2023 12:15
 Operator : JC/MD
 Sample : 03491-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-05-167-187-062923

Quant Time: Jul 01 01:19:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	177889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	301281	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121064	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	111099	43.570	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.140%
35) Dibromofluoromethane	5.385	113	92579	46.603	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.200%
50) Toluene-d8	8.647	98	352185	47.466	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.940%
62) 4-Bromofluorobenzene	11.079	95	125502	48.021	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.040%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	201620	99.730	ug/l	92
16) Acetone	2.392	43	3133	2.875	ug/l	91
21) trans-1,2-Dichloroethene	3.093	96	20329	9.423	ug/l	94
24) 1,1-Dichloroethane	3.611	63	27920	7.196	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	152693	63.112	ug/l	84
40) Benzene	6.037	78	3230	0.390	ug/l	95
42) 1,2-Dichloroethane	6.092	62	16807	5.342	ug/l	99
44) Trichloroethene	7.123	130	143276	62.271	ug/l	100
49) 1,4-Dioxane	7.702	88	5468	102.746	ug/l #	89
52) Toluene	8.720	92	114127	21.770	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063023\
Data File : VX036387.D
Acq On : 30 Jun 2023 12:15
Operator : JC/MD
Sample : 03491-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-05-167-187-062923

Quant Time: Jul 01 01:19:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

