

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063023\
 Data File : VX036388.D
 Acq On : 30 Jun 2023 12:39
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 01 01:20:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	189849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	313511	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	277423	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129867	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	119097	43.764	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.520%
35) Dibromofluoromethane	5.391	113	98285	47.545	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.080%
50) Toluene-d8	8.647	98	374610	48.518	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.040%
62) 4-Bromofluorobenzene	11.079	95	134876	49.594	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063023\
Data File : VX036388.D
Acq On : 30 Jun 2023 12:39
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 01 01:20:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

