

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063023\
 Data File : VX036390.D
 Acq On : 30 Jun 2023 13:26
 Operator : JC/MD
 Sample : 03494-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 01 01:21:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	179742	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	301203	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	265003	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120420	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	114066	44.272	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.540%
35) Dibromofluoromethane	5.385	113	93617	47.137	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.280%
50) Toluene-d8	8.647	98	354013	47.724	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.440%
62) 4-Bromofluorobenzene	11.079	95	123909	47.423	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063023\
Data File : VX036390.D
Acq On : 30 Jun 2023 13:26
Operator : JC/MD
Sample : 03494-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 01 01:21:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

