

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063023\
 Data File : VX036391.D
 Acq On : 30 Jun 2023 13:50
 Operator : JC/MD
 Sample : 03494-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 01 01:22:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	179166	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	298894	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	267874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119779	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	113174	44.067	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.140%
35) Dibromofluoromethane	5.385	113	91729	46.543	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.080%
50) Toluene-d8	8.647	98	355859	48.344	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.680%
62) 4-Bromofluorobenzene	11.079	95	123020	47.447	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063023\
Data File : VX036391.D
Acq On : 30 Jun 2023 13:50
Operator : JC/MD
Sample : 03494-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 01 01:22:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

