

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063025\
 Data File : VX046830.D
 Acq On : 30 Jun 2025 10:00
 Operator : JC/MD
 Sample : VX0630WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0630WBL01

Quant Time: Jul 01 05:16:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	114774	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	254959	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	328591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	181976	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	93182	58.696	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.400%
35) Dibromofluoromethane	5.397	113	85976	50.967	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.940%
50) Toluene-d8	8.647	98	330491	54.531	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.060%
62) 4-Bromofluorobenzene	11.079	95	178706	79.102	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	158.200%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063025\
Data File : VX046830.D
Acq On : 30 Jun 2025 10:00
Operator : JC/MD
Sample : VX0630WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0630WBL01

Quant Time: Jul 01 05:16:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

