

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063025\
 Data File : VX046835.D
 Acq On : 30 Jun 2025 12:39
 Operator : JC/MD
 Sample : Q2398-06ME
 Misc : 4.60g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ARS20-0036ME

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/30/2025
 Supervised By : Mahesh Dadoda 07/01/2025

Quant Time: Jun 30 12:55:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	205270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	346748	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	317038	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	155853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	132743	46.753	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.500%		
35) Dibromofluoromethane	5.391	113	109536	47.745	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.480%		
50) Toluene-d8	8.647	98	416761	50.562	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.120%		
62) 4-Bromofluorobenzene	11.079	95	156468	50.925	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	101.860%		
Target Compounds						Qvalue
16) Acetone	2.398	43	63010	66.549	ug/l	98
20) Methylene Chloride	2.794	84	690m	0.262	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063025\
Data File : VX046835.D
Acq On : 30 Jun 2025 12:39
Operator : JC/MD
Sample : Q2398-06ME
Misc : 4.60g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ARS20-0036ME

Quant Time: Jun 30 12:55:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/30/2025
Supervised By : Mahesh Dadoda 07/01/2025

