

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063025\
 Data File : VX046838.D
 Acq On : 30 Jun 2025 13:48
 Operator : JC/MD
 Sample : Q2371-04DL 800X
 Misc : 6.82g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RPXY42025DL

Quant Time: Jul 01 05:30:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	267442	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	441134	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	397169	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	196471	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	161055	43.538	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.080%
35) Dibromofluoromethane	5.397	113	139226	47.702	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.400%
50) Toluene-d8	8.647	98	519970	49.586	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.180%
62) 4-Bromofluorobenzene	11.079	95	198922	50.890	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.780%
Target Compounds						
					Qvalue	
52) Toluene	8.726	92	2397	0.310	ug/l	97
67) Ethyl Benzene	10.195	91	211160	13.784	ug/l	99
68) m/p-Xylenes	10.299	106	283556	49.678	ug/l	99
69) o-Xylene	10.640	106	69746	13.038	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX063025\
Data File : VX046838.D
Acq On : 30 Jun 2025 13:48
Operator : JC/MD
Sample : Q2371-04DL 800X
Misc : 6.82g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RPXY42025DL

Quant Time: Jul 01 05:30:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

