

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070119\
 Data File : VX010586.D
 Acq On : 01 Jul 2019 13:24
 Operator : JC/SP
 Sample : K3565-01DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0568DL

Quant Time: Jul 02 10:13:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	247471	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	374458	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	344859	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	167644	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118357	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
35) Dibromofluoromethane	5.49	113	112665	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
50) Toluene-d8	8.71	98	447432	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
62) 4-Bromofluorobenzene	11.13	95	153750	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
Target Compounds						
40) Benzene	6.13	78	74640	8.051	ug/l	98
67) Ethyl Benzene	10.25	91	307041	25.977	ug/l	100
68) m/p-Xylenes	10.36	106	125643	27.434	ug/l	99
84) 1,2,4-Trimethylbenzene	11.80	105	130291	13.843	ug/l	97
95) Naphthalene	13.83	128	61259	5.222	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070119\
Data File : VX010586.D
Acq On : 01 Jul 2019 13:24
Operator : JC/SP
Sample : K3565-01DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0568DL

Quant Time: Jul 02 10:13:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

