

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070119\
 Data File : VX010595.D
 Acq On : 01 Jul 2019 16:54
 Operator : JC/SP
 Sample : K3557-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-EW03-TRENCH

Quant Time: Jul 02 07:06:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	242223	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	375811	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	352468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	161286	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117774	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
35) Dibromofluoromethane	5.50	113	113494	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	8.71	98	453328	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
62) 4-Bromofluorobenzene	11.13	95	151563	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
Target Compounds						
16) Acetone	2.45	43	11801	10.368	ug/l	94
20) Methylene Chloride	2.86	84	13533	5.778	ug/l	97
43) Isopropyl Acetate	6.45	43	9619	1.892	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070119\
Data File : VX010595.D
Acq On : 01 Jul 2019 16:54
Operator : JC/SP
Sample : K3557-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-EW03-TRENCH

Quant Time: Jul 02 07:06:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

