

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070119\
 Data File : VX010597.D
 Acq On : 01 Jul 2019 17:41
 Operator : JC/SP
 Sample : K3585-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RM-WATER-COMP

Quant Time: Jul 02 07:10:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	248908	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	385914	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	350098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	160471	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	121516	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
35) Dibromofluoromethane	5.50	113	117006	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
50) Toluene-d8	8.71	98	454360	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	11.13	95	152960	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
Target Compounds						
11) Tert butyl alcohol	3.05	59	3744	6.131	ug/l #	82
16) Acetone	2.44	43	29491	25.214	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070119\
Data File : VX010597.D
Acq On : 01 Jul 2019 17:41
Operator : JC/SP
Sample : K3585-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RM-WATER-COMP

Quant Time: Jul 02 07:10:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

