

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070119\
 Data File : VX010599.D
 Acq On : 01 Jul 2019 18:28
 Operator : JC/SP
 Sample : K3605-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 QNWP8-TRIP-BLANK

Quant Time: Jul 02 07:14:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	249060	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	385250	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	344575	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	159564	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119624	51.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.46%	
35) Dibromofluoromethane	5.50	113	115309	48.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.84%	
50) Toluene-d8	8.71	98	451871	50.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.22%	
62) 4-Bromofluorobenzene	11.13	95	151421	46.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.02%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070119\
Data File : VX010599.D
Acq On : 01 Jul 2019 18:28
Operator : JC/SP
Sample : K3605-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
QNWP8-TRIP-BLANK

Quant Time: Jul 02 07:14:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

