

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070119\
 Data File : VX010600.D
 Acq On : 01 Jul 2019 18:51
 Operator : JC/SP
 Sample : K3604-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VO-BLANK

Quant Time: Jul 02 07:16:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	253735	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	392915	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	358799	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	165203	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	121697	51.15	ug/l	0.00
Spiked Amount						
			Recovery	=	102.30%	
35) Dibromofluoromethane	5.50	113	117356	48.82	ug/l	0.00
Spiked Amount						
			Recovery	=	97.64%	
50) Toluene-d8	8.71	98	465051	50.56	ug/l	0.00
Spiked Amount						
			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.13	95	154952	46.67	ug/l	0.00
Spiked Amount						
			Recovery	=	93.34%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070119\
Data File : VX010600.D
Acq On : 01 Jul 2019 18:51
Operator : JC/SP
Sample : K3604-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VO-BLANK

Quant Time: Jul 02 07:16:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

