

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070120\
 Data File : VX017091.D
 Acq On : 01 Jul 2020 12:36
 Operator : JC/SP
 Sample : L3129-01 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1446

Quant Time: Jul 02 02:09:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	224275	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	375715	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	388396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	195232	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	148873	53.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.68%	
35) Dibromofluoromethane	5.47	113	126647	53.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.58%	
50) Toluene-d8	8.70	98	465527	52.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.94%	
62) 4-Bromofluorobenzene	11.12	95	191618	55.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.56%	
Target Compounds						
16) Acetone	2.42	43	3918	3.788	ug/l #	Qvalue 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070120\
Data File : VX017091.D
Acq On : 01 Jul 2020 12:36
Operator : JC/SP
Sample : L3129-01 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1446

Quant Time: Jul 02 02:09:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

