

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070120\
 Data File : VX017094.D
 Acq On : 01 Jul 2020 13:44
 Operator : JC/SP
 Sample : L3129-11 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1308

Quant Time: Jul 02 02:15:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	219949	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	372360	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	382297	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	203649	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	146846	54.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.32%	
35) Dibromofluoromethane	5.47	113	123592	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
50) Toluene-d8	8.70	98	458797	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
62) 4-Bromofluorobenzene	11.12	95	188905	55.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.96%	
Target Compounds						
11) Tert butyl alcohol	3.01	59	2714	5.247	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070120\
Data File : VX017094.D
Acq On : 01 Jul 2020 13:44
Operator : JC/SP
Sample : L3129-11 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1308

Quant Time: Jul 02 02:15:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

